

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091918\
 Data File : VX004697.D
 Acq On : 19 Sep 2018 11:17
 Operator : JC/MD
 Sample : VX0919MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0919MBL01

Quant Time: Sep 19 16:57:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 02:17:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	232503	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	322281	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	297960	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	164348	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	144531	48.21	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	96.42%	
35) Dibromofluoromethane	5.50	113	134317	45.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.08%	
50) Toluene-d8	8.71	98	443175	44.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.62%	
62) 4-Bromofluorobenzene	11.14	95	153692	43.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.06%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091918\
Data File : VX004697.D
Acq On : 19 Sep 2018 11:17
Operator : JC/MD
Sample : VX0919MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0919MBL01

Quant Time: Sep 19 16:57:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091118W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 02:17:23 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX004697.D

Acq: 19 Sep 2018 11:17

Instrument :

MSVOA_X

ClientSampleId :


VX0919MBL01

Tot Ion:168 Resp: 232503

Ion Ratio Lower Upper




168 100

99 44.5 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 48.21 ug/l

RT: 6.06 min Scan# 804

Delta R.T. -0.01 min

Lab File: VX004697.D


Acq: 19 Sep 2018 11:17

Tgt Ion: 65 Resp: 144531

Ion Ratio Lower Upper



65 100


67 57.1 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 937
 Delta R.T. 0.01 min
 Lab File: VX004697.D
 Acq: 19 Sep 2018 11:17

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0919MBL01

Tot Ion:114 Resp: 322281
 Ion Ratio Lower Upper
 114 100
 63 16.3 0.0 39.0
 88 13.9 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 45.54 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. 0.00 min
 Lab File: VX004697.D
 Acq: 19 Sep 2018 11:17

Tgt Ion:113 Resp: 134317
 Ion Ratio Lower Upper
 113 100
 111 102.4 82.4 123.6
 192 24.5 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 44.81 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.01 min

Lab File: VX004697.D

Acq: 19 Sep 2018 11:17

Instrument :

MSVOA_X

ClientSampleId :


VX0919MBL01

Tot Ion: 98 Resp: 443175

Ion Ratio Lower Upper

98 100

100 63.6 52.9 79.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71



300000

200000

100000

0

Time-->



#62

4-Bromofluorobenzene

Concen: 43.53 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.00 min

Lab File: VX004697.D

Acq: 19 Sep 2018 11:17


Tgt Ion: 95 Resp: 153692

Ion Ratio Lower Upper

95 100

174 95.9 0.0 185.2

176 92.0 0.0 178.6



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.14


150000


100000

50000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX004697.D
Acq: 19 Sep 2018 11:17

Instrument :
MSVOA_X
ClientSampleId :
VX0919MBL01

Tot Ion:117 Resp: 297960
Ion Ratio Lower Upper
117 100
82 52.6 47.8 71.6
119 32.9 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX004697.D
Acq: 19 Sep 2018 11:17

Tgt Ion:152 Resp: 164348
Ion Ratio Lower Upper
152 100
115 52.8 39.0 117.0
150 154.8 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

