

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091919\
 Data File : VX012523.D
 Acq On : 19 Sep 2019 13:14
 Operator : JC/SP
 Sample : K4888-01
 Misc : 6.64g/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 979-S-1-(18-19)

Quant Time: Sep 20 05:13:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	178962	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	278928	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	237019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	107887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	114447	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
35) Dibromofluoromethane	5.49	113	77533	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
50) Toluene-d8	8.71	98	323030	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.14	95	119663	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
Target Compounds						
44) Trichloroethene	7.21	130	1843	1.300	ug/l	92
64) Tetrachloroethene	9.33	164	371649	342.183	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091919\
Data File : VX012523.D
Acq On : 19 Sep 2019 13:14
Operator : JC/SP
Sample : K4888-01
Misc : 6.64g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
979-S-1-(18-19)

Quant Time: Sep 20 05:13:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

