

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091919\
 Data File : VX012532.D
 Acq On : 19 Sep 2019 16:44
 Operator : JC/SP
 Sample : K4883-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW6-R2-1

Quant Time: Sep 26 08:58:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	181386	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	289038	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	245156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101340	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	118498	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery =	91.16%		
35) Dibromofluoromethane	5.49	113	84198	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery =	96.26%		
50) Toluene-d8	8.72	98	334573	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery =	103.34%		
62) 4-Bromofluorobenzene	11.14	95	123192	45.82	ug/l	0.00
Spiked Amount 50.000			Recovery =	91.64%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	2699599	2504.720	ug/l	95
12) 1,1-Dichloroethene	2.37	96	335640	294.322	ug/l	98
16) Acetone	2.44	43	13547	3.558	ug/l	97
21) trans-1,2-Dichloroethene	3.16	96	227929	192.474	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	20001403	11306.481	ug/l	83
30) Chloroform	5.21	83	3343	0.979	ug/l	97
31) Cyclohexane	5.57	56	24978	13.912	ug/l	99
39) Methylcyclohexane	7.46	83	24387	14.281	ug/l	94
40) Benzene	6.14	78	74130	13.271	ug/l	96
44) Trichloroethene	7.22	130	8576545	5836.687	ug/l	94
52) Toluene	8.78	92	16256	4.516	ug/l	94
64) Tetrachloroethene	9.36	164	12256452	10910.161	ug/l	94
67) Ethyl Benzene	10.25	91	9152	1.316	ug/l	95
68) m/p-Xylenes	10.36	106	8253	3.263	ug/l	91
69) o-Xylene	10.70	106	5126	1.971	ug/l	97
73) Isopropylbenzene	11.02	105	3028	0.476	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	3144	0.565	ug/l	100
88) 1,4-Dichlorobenzene	12.09	146	20178	6.734	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091919\
Data File : VX012532.D
Acq On : 19 Sep 2019 16:44
Operator : JC/SP
Sample : K4883-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW6-R2-1

Quant Time: Sep 26 08:58:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

