

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091919\
 Data File : VX012534.D
 Acq On : 19 Sep 2019 17:30
 Operator : JC/SP
 Sample : K4888-01DL 10X
 Misc : 6.64g/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 979-S-1-(18-19)DL

Quant Time: Sep 20 05:56:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	185237	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	289696	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244307	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92997	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	124313	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
35) Dibromofluoromethane	5.49	113	87865	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	8.71	98	338378	52.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.28%	
62) 4-Bromofluorobenzene	11.14	95	120895	44.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.72%	
Target Compounds						
4) Vinyl Chloride	1.40	62	4276	3.885	ug/l	99
27) cis-1,2-Dichloroethene	4.59	96	38038	21.055	ug/l	84
44) Trichloroethene	7.21	130	26341	17.885	ug/l	95
55) 1,1,2-Trichloroethane	9.33	97	1062	0.571	ug/l #	21
64) Tetrachloroethene	9.33	164	101628	90.779	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091919\
Data File : VX012534.D
Acq On : 19 Sep 2019 17:30
Operator : JC/SP
Sample : K4888-01DL 10X
Misc : 6.64g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
979-S-1-(18-19)DL

Quant Time: Sep 20 05:56:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

