

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX091919\
 Data File : VX012537.D
 Acq On : 19 Sep 2019 18:41
 Operator : JC/SP
 Sample : K4858-25DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20190911DL

Quant Time: Sep 20 06:53:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	185761	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	299449	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	248941	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	94009	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	128415	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
35) Dibromofluoromethane	5.49	113	90026	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
50) Toluene-d8	8.71	98	340820	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
62) 4-Bromofluorobenzene	11.14	95	119070	42.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.50%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	4666	3.187	ug/l	92
27) cis-1,2-Dichloroethene	4.58	96	6044	3.336	ug/l	84
44) Trichloroethene	7.21	130	56618	37.191	ug/l	96
64) Tetrachloroethene	9.33	164	11171	9.793	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX091919\
Data File : VX012537.D
Acq On : 19 Sep 2019 18:41
Operator : JC/SP
Sample : K4858-25DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20190911DL

Quant Time: Sep 20 06:53:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

