


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX091919\
Data File : VX012516.D
Acq On : 19 Sep 2019 09:58
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Title : SW846 8260
Last Update : Tue Sep 17 15:27:10 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1671

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	7670	PASS
75	95	30	60	51.0	19804	PASS
95	95	100	100	100.0	38808	PASS
96	95	5	9	7.2	2798	PASS
173	174	0.00	2	0.2	58	PASS
174	95	50	100	64.5	25049	PASS
175	174	5	9	7.6	1908	PASS
176	174	95	101	97.8	24504	PASS
177	176	5	9	6.4	1558	PASS