

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031466.D
 Acq On : 19 Sep 2022 17:39
 Operator : JC/MD
 Sample : N4509-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1-SO_116D-20220902

Quant Time: Sep 20 09:49:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	119508	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110503	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88413	66.426	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 132.860%#	
35) Dibromofluoromethane	5.391	113	73000	59.232	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 118.460%	
50) Toluene-d8	8.653	98	268462	56.384	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 112.760%	
62) 4-Bromofluorobenzene	11.079	95	83214	51.596	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.200%	
Target Compounds						Qvalue
37) Ethyl Acetate	4.721	43	11318	4.994	ug/l	99
43) Isopropyl Acetate	6.342	43	42064	12.077	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031466.D
Acq On : 19 Sep 2022 17:39
Operator : JC/MD
Sample : N4509-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1-SO_116D-20220902

Quant Time: Sep 20 09:49:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

