

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031468.D
 Acq On : 19 Sep 2022 18:26
 Operator : JC/MD
 Sample : N4558-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-TB-20220906RE

Quant Time: Sep 20 09:49:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	72119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	117742	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77841	58.937	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.880%			
35) Dibromofluoromethane	5.391	113	65292	53.704	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.400%			
50) Toluene-d8	8.653	98	244719	52.168	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.340%			
62) 4-Bromofluorobenzene	11.079	95	73561	46.295	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.580%			
Target Compounds						
30) Chloroform	5.099	83	7495	3.050	ug/l	99
47) Bromodichloromethane	7.830	83	979	0.482	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031468.D
Acq On : 19 Sep 2022 18:26
Operator : JC/MD
Sample : N4558-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-TB-20220906RE

Quant Time: Sep 20 09:49:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

