

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031473.D
 Acq On : 19 Sep 2022 20:23
 Operator : JC/MD
 Sample : N4558-09RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-278-280RE

Quant Time: Sep 20 09:50:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67798	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	113348	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104699	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46284	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74672	60.163	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.320%
35) Dibromofluoromethane	5.391	113	62693	53.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.120%
50) Toluene-d8	8.653	98	238205	52.748	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.500%
62) 4-Bromofluorobenzene	11.085	95	72977	47.708	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031473.D
Acq On : 19 Sep 2022 20:23
Operator : JC/MD
Sample : N4558-09RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-278-280RE

Quant Time: Sep 20 09:50:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

