

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031476.D
 Acq On : 19 Sep 2022 21:34
 Operator : JC/MD
 Sample : N4688-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-DUP-20220914RE

Quant Time: Sep 20 09:50:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	110960	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72025	58.298	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.600%
35) Dibromofluoromethane	5.385	113	60860	53.110	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.220%
50) Toluene-d8	8.653	98	227370	51.432	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.860%
62) 4-Bromofluorobenzene	11.085	95	68868	45.991	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031476.D
Acq On : 19 Sep 2022 21:34
Operator : JC/MD
Sample : N4688-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-DUP-20220914RE

Quant Time: Sep 20 09:50:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

