

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031479.D
 Acq On : 19 Sep 2022 22:44
 Operator : JC/MD
 Sample : N4688-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-478-480RE

Quant Time: Sep 20 09:51:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	110672	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72460	58.728	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.460%
35) Dibromofluoromethane	5.385	113	60399	52.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.680%
50) Toluene-d8	8.653	98	225501	51.142	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.280%
62) 4-Bromofluorobenzene	11.085	95	67143	44.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.920%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	607	0.627	ug/l	86
9) 1,1,2-Trichlorotrifluo...	2.331	101	3143	2.897	ug/l	90
12) 1,1-Dichloroethene	2.312	96	511	0.487	ug/l #	58
27) cis-1,2-Dichloroethene	4.501	96	1258	0.881	ug/l	66
44) Trichloroethene	7.129	130	52778	35.468	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031479.D
Acq On : 19 Sep 2022 22:44
Operator : JC/MD
Sample : N4688-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-478-480RE

Quant Time: Sep 20 09:51:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

