

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
 Data File : VX031481.D
 Acq On : 19 Sep 2022 23:31
 Operator : JC/MD
 Sample : N4653-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CAR-1948

Quant Time: Sep 20 09:51:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	111738	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75004	60.693	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.380%	
35) Dibromofluoromethane	5.391	113	64054	55.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.080%	
50) Toluene-d8	8.653	98	245628	55.176	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 110.360%	
62) 4-Bromofluorobenzene	11.085	95	78180	51.846	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	9153	5.324	ug/l	97
20) Methylene Chloride	2.788	84	5511	2.774	ug/l	92
37) Ethyl Acetate	4.727	43	10010	4.724	ug/l	98
43) Isopropyl Acetate	6.342	43	44277	13.596	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091922\
Data File : VX031481.D
Acq On : 19 Sep 2022 23:31
Operator : JC/MD
Sample : N4653-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CAR-1948

Quant Time: Sep 20 09:51:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

