

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037793.D
 Acq On : 19 Sep 2023 10:12
 Operator : JC/MD
 Sample : VX0919MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0919MBL01

Quant Time: Sep 20 00:57:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	249885	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	415081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	385482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	189192	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	150272	45.771	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.540%
35) Dibromofluoromethane	5.385	113	128357	46.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.440%
50) Toluene-d8	8.647	98	499531	48.407	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.820%
62) 4-Bromofluorobenzene	11.079	95	179433	47.850	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037793.D
Acq On : 19 Sep 2023 10:12
Operator : JC/MD
Sample : VX0919MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0919MBL01

Quant Time: Sep 20 00:57:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

