

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037798.D
 Acq On : 19 Sep 2023 12:11
 Operator : JC/MD
 Sample : 04479-02 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1027

Quant Time: Sep 20 01:00:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	240842	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	401915	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	379788	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	145751	46.061	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.120%
35) Dibromofluoromethane	5.391	113	124220	46.192	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	492517	49.290	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.580%
62) 4-Bromofluorobenzene	11.079	95	179539	49.447	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.900%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	3226	2.119	ug/l	# 82
20) Methylene Chloride	2.794	84	1733	0.591	ug/l	90
25) 2-Butanone	4.574	43	8803	4.098	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037798.D
Acq On : 19 Sep 2023 12:11
Operator : JC/MD
Sample : 04479-02 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1027

Quant Time: Sep 20 01:00:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

