

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037800.D
 Acq On : 19 Sep 2023 12:57
 Operator : JC/MD
 Sample : 04465-02 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWC

Quant Time: Sep 20 01:01:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	246838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	414755	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	400468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	205597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	148592	45.818	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	91.640%	
35) Dibromofluoromethane	5.391	113	127352	45.890	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	91.780%	
50) Toluene-d8	8.647	98	507492	49.217	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	98.440%	
62) 4-Bromofluorobenzene	11.079	95	192664	51.419	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	102.840%	
Target Compounds						
20) Methylene Chloride	2.794	84	2387	0.803	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037800.D
Acq On : 19 Sep 2023 12:57
Operator : JC/MD
Sample : 04465-02 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWC

Quant Time: Sep 20 01:01:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

