

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037803.D
 Acq On : 19 Sep 2023 14:07
 Operator : JC/MD
 Sample : 04491-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-9-15

Quant Time: Sep 20 01:02:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	224204	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	378190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	353904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139223	47.263	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.520%
35) Dibromofluoromethane	5.391	113	117193	46.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.620%
50) Toluene-d8	8.653	98	458922	48.809	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.079	95	162912	47.682	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.360%
Target Compounds						
16) Acetone	2.380	43	7130	5.032	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037803.D
Acq On : 19 Sep 2023 14:07
Operator : JC/MD
Sample : 04491-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-9-15

Quant Time: Sep 20 01:02:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

