

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037805.D
 Acq On : 19 Sep 2023 14:53
 Operator : JC/MD
 Sample : 04414-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-08B-18-091223

Quant Time: Sep 20 01:03:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	219585	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	371867	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	356741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	183913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138490	48.003	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.000%
35) Dibromofluoromethane	5.391	113	115074	46.248	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	452387	48.932	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	165824	49.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.720%
Target Compounds						
16) Acetone	2.380	43	2791	2.011	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037805.D
Acq On : 19 Sep 2023 14:53
Operator : JC/MD
Sample : 04414-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-08B-18-091223

Quant Time: Sep 20 01:03:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

