

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
 Data File : VX037813.D
 Acq On : 19 Sep 2023 17:59
 Operator : JC/MD
 Sample : PB155628TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB155628TB

Quant Time: Sep 20 01:06:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	366918	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372524	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	193747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135345	48.057	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.120%
35) Dibromofluoromethane	5.391	113	114221	46.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	8.647	98	457307	50.132	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	177086	53.423	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.840%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	23861	17.612	ug/l	92
20) Methylene Chloride	2.794	84	13372	5.316	ug/l #	88
25) 2-Butanone	4.568	43	14824	7.754	ug/l	97
43) Isopropyl Acetate	6.342	43	153902	25.887	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091923\
Data File : VX037813.D
Acq On : 19 Sep 2023 17:59
Operator : JC/MD
Sample : PB155628TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB155628TB

Quant Time: Sep 20 01:06:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

