

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047687.D
 Acq On : 19 Sep 2025 19:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 20 00:37:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	154405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	278359	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	248998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	120684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	130953	53.282	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.560%
35) Dibromofluoromethane	5.373	113	100288	53.721	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.440%
50) Toluene-d8	8.635	98	344346	53.308	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.620%
62) 4-Bromofluorobenzene	11.061	95	129494	52.822	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.640%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.179	85	101670	63.588	ug/l	98
3) Chloromethane	1.300	50	123827	58.927	ug/l	99
4) Vinyl Chloride	1.380	62	122810	59.703	ug/l	97
5) Bromomethane	1.617	94	77609	59.499	ug/l	100
6) Chloroethane	1.697	64	77528	56.391	ug/l	97
7) Trichlorofluoromethane	1.898	101	168856	55.285	ug/l	99
8) Diethyl Ether	2.136	74	68004	54.429	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.337	101	99678	55.814	ug/l	97
10) Methyl Iodide	2.459	142	146871	52.731	ug/l	100
11) Tert butyl alcohol	2.940	59	58211	212.661	ug/l	100
12) 1,1-Dichloroethene	2.325	96	100925	55.015	ug/l	95
13) Acrolein	2.239	56	68460	271.404	ug/l	98
14) Allyl chloride	2.666	41	195491	51.642	ug/l	98
15) Acrylonitrile	3.056	53	276208	272.421	ug/l	100
16) Acetone	2.373	43	211924	198.313	ug/l	96
17) Carbon Disulfide	2.520	76	289670	57.679	ug/l	100
18) Methyl Acetate	2.697	43	127073	53.432	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	351150	52.436	ug/l	99
20) Methylene Chloride	2.788	84	122715	54.250	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	109388	55.385	ug/l	97
22) Diisopropyl ether	3.751	45	400135	54.541	ug/l #	83
23) Vinyl Acetate	3.715	43	1611770	274.498	ug/l	100
24) 1,1-Dichloroethane	3.605	63	212751	54.560	ug/l	99
25) 2-Butanone	4.538	43	329510	247.197	ug/l	99
26) 2,2-Dichloropropane	4.471	77	145629	44.928	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	129834	53.677	ug/l	98
28) Bromochloromethane	4.885	49	91361	53.668	ug/l	99
29) Tetrahydrofuran	4.989	42	208846	260.441	ug/l	100
30) Chloroform	5.080	83	214531	54.443	ug/l	96
31) Cyclohexane	5.464	56	173386	53.387	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	178728	53.474	ug/l	100
36) 1,1-Dichloropropene	5.684	75	140108	51.362	ug/l	99
37) Ethyl Acetate	4.696	43	147497	50.499	ug/l	100
38) Carbon Tetrachloride	5.666	117	150656	50.831	ug/l	95
39) Methylcyclohexane	7.366	83	162143	52.363	ug/l	99
40) Benzene	6.025	78	445760	53.802	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047687.D
 Acq On : 19 Sep 2025 19:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 20 00:37:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.903	41	79917	53.000	ug/l	99
42) 1,2-Dichloroethane	6.074	62	163773	51.951	ug/l	98
43) Isopropyl Acetate	6.324	43	246265	51.410	ug/l	99
44) Trichloroethene	7.117	130	112580	56.204	ug/l	93
45) 1,2-Dichloropropane	7.415	63	114044	53.755	ug/l	99
46) Dibromomethane	7.568	93	81845	53.014	ug/l	97
47) Bromodichloromethane	7.805	83	169676	53.260	ug/l	97
48) Methyl methacrylate	7.677	41	124085	52.004	ug/l	98
49) 1,4-Dioxane	7.647	88	26594	1102.850	ug/l	97
51) 4-Methyl-2-Pentanone	8.555	43	702538	264.517	ug/l	99
52) Toluene	8.702	92	272421	53.368	ug/l	99
53) t-1,3-Dichloropropene	8.964	75	167442	51.527	ug/l	99
54) cis-1,3-Dichloropropene	8.354	75	179970	51.815	ug/l	99
55) 1,1,2-Trichloroethane	9.134	97	104952	53.322	ug/l	99
56) Ethyl methacrylate	9.104	69	165642	53.586	ug/l	98
57) 1,3-Dichloropropane	9.293	76	183018	54.129	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.226	63	384468	259.650	ug/l	100
59) 2-Hexanone	9.415	43	481662	252.510	ug/l	99
60) Dibromochloromethane	9.506	129	121253	53.465	ug/l	99
61) 1,2-Dibromoethane	9.592	107	110255	54.996	ug/l	99
64) Tetrachloroethene	9.262	164	84740	52.205	ug/l	95
65) Chlorobenzene	10.061	112	298639	52.793	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.146	131	101896	52.201	ug/l	97
67) Ethyl Benzene	10.177	91	517118	52.982	ug/l	99
68) m/p-Xylenes	10.287	106	381354	104.973	ug/l	99
69) o-Xylene	10.628	106	187399	53.332	ug/l	99
70) Styrene	10.640	104	327913	53.259	ug/l	99
71) Bromoform	10.780	173	75620	51.882	ug/l #	98
73) Isopropylbenzene	10.945	105	475596	51.835	ug/l	100
74) N-amyl acetate	10.829	43	216669	51.785	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.195	83	145518	52.421	ug/l	98
76) 1,2,3-Trichloropropane	11.219	75	158337	64.800	ug/l	89
77) Bromobenzene	11.183	156	117390	51.947	ug/l	99
78) n-propylbenzene	11.286	91	565595	52.147	ug/l	100
79) 2-Chlorotoluene	11.347	91	341195	51.361	ug/l	100
80) 1,3,5-Trimethylbenzene	11.433	105	388702	51.564	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.000	75	43877	45.792	ug/l	97
82) 4-Chlorotoluene	11.439	91	401172	51.132	ug/l	98
83) tert-Butylbenzene	11.695	119	391982	50.797	ug/l	99
84) 1,2,4-Trimethylbenzene	11.732	105	397498	52.011	ug/l	100
85) sec-Butylbenzene	11.872	105	467967	51.055	ug/l	100
86) p-Isopropyltoluene	11.988	119	392640	50.599	ug/l	99
87) 1,3-Dichlorobenzene	11.951	146	215979	51.456	ug/l	99
88) 1,4-Dichlorobenzene	12.024	146	218234	50.659	ug/l	100
89) n-Butylbenzene	12.311	91	365418	50.892	ug/l	98
90) Hexachloroethane	12.518	117	68460	50.250	ug/l	98
91) 1,2-Dichlorobenzene	12.317	146	206378	51.610	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.920	75	26906	47.873	ug/l	95
93) 1,2,4-Trichlorobenzene	13.567	180	132061	50.815	ug/l	99
94) Hexachlorobutadiene	13.707	225	40751	46.163	ug/l	98
95) Naphthalene	13.756	128	395649	49.995	ug/l	100
96) 1,2,3-Trichlorobenzene	13.938	180	118876	49.173	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047687.D
Acq On : 19 Sep 2025 19:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Sep 20 00:37:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 20 00:37:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 154405
Ion Ratio Lower Upper
168 100
99 61.0 48.8 73.2



#2
Dichlorodifluoromethane
Concen: 63.588 ug/l
RT: 1.179 min Scan# 15
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 85 Resp: 101670
Ion Ratio Lower Upper
85 100
87 32.5 15.8 47.3









#3
Chloromethane
Concen: 58.927 ug/l
RT: 1.300 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 123827
Ion Ratio Lower Upper
50 100
52 31.8 26.1 39.1



#4
Vinyl Chloride
Concen: 59.703 ug/l
RT: 1.380 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03


Tgt Ion: 62 Resp: 122810
Ion Ratio Lower Upper
62 100
64 32.7 24.9 37.3








#5
 Bromomethane
 Concen: 59.499 ug/l
 RT: 1.617 min Scan# 81
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion: 94 Resp: 77609
 Ion Ratio Lower Upper
 94 100
 96 95.8 76.6 115.0



#6
 Chloroethane
 Concen: 56.391 ug/l
 RT: 1.697 min Scan# 100
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03


Tgt Ion: 64 Resp: 77528
 Ion Ratio Lower Upper
 64 100
 66 31.1 26.1 39.1








#7
 Trichlorofluoromethane
 Concen: 55.285 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03

Instrument :
 MSVOA_X
 ClientSampleId :






Tgt Ion:101 Resp: 168856
 Ion Ratio Lower Upper
 101 100
 103 64.8 51.4 77.2



#8
 Diethyl Ether
 Concen: 54.429 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03

Tgt Ion: 74 Resp: 68004
 Ion Ratio Lower Upper
 74 100
 45 103.2 52.3 156.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 55.814 ug/l
RT: 2.337 min Scan# 205
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 99678
Ion Ratio Lower Upper
101 100
85 43.2 35.5 53.3
151 71.7 59.4 89.0



#10
Methyl Iodide
Concen: 52.731 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:142 Resp: 146871
Ion Ratio Lower Upper
142 100
127 46.7 37.2 55.8
141 14.4 11.4 17.0





#11

Tert butyl alcohol

Concen: 212.661 ug/l

RT: 2.940 min Scan# 304

Delta R.T. -0.000 min


Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 59 Resp: 58211

Ion Ratio Lower Upper

59 100

57 10.5 8.3 12.5



#12

1,1-Dichloroethene

Concen: 55.015 ug/l

RT: 2.325 min Scan# 203

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 96 Resp: 100925


Ion Ratio Lower Upper

96 100

61 169.9 131.0 196.6

98 65.2 47.7 71.5





#13

Acrolein

Concen: 271.404 ug/l

RT: 2.239 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 68460

Ion Ratio Lower Upper

56 100

55 70.2 57.2 85.8



#14

Allyl chloride

Concen: 51.642 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 41 Resp: 195491


Ion Ratio Lower Upper

41 100

39 66.2 54.7 82.1


76 32.4 25.7 38.5








#15
Acrylonitrile
Concen: 272.421 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 276208
Ion Ratio Lower Upper
53 100
52 82.9 66.4 99.6
51 36.7 29.5 44.3



#16
Acetone
Concen: 198.313 ug/l
RT: 2.373 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03


Tgt Ion: 43 Resp: 211924
Ion Ratio Lower Upper
43 100
58 31.2 23.4 35.0








#17
Carbon Disulfide
Concen: 57.679 ug/l
RT: 2.520 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 76 Resp: 289670
Ion Ratio Lower Upper
76 100
78 8.6 6.8 10.2



#18
Methyl Acetate
Concen: 53.432 ug/l
RT: 2.697 min Scan# 264
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 43 Resp: 127073
Ion Ratio Lower Upper
43 100
74 21.7 17.4 26.2





#19

Methyl tert-butyl Ether

Concen: 52.436 ug/l

RT: 3.111 min Scan# 311

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 73 Resp: 351150

Ion Ratio Lower Upper

73 100

57 22.2 17.4 26.2



#20

Methylene Chloride

Concen: 54.250 ug/l

RT: 2.788 min Scan# 279

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Tgt Ion: 84 Resp: 122715




Ion Ratio Lower Upper


84 100

49 131.7 108.6 162.8

51 40.1 32.6 48.8

86 62.6 52.0 78.0









#21
trans-1,2-Dichloroethene
Concen: 55.385 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 109388
Ion Ratio Lower Upper
96 100
61 142.7 116.2 174.2
98 65.1 49.3 73.9



#22
Diisopropyl ether
Concen: 54.541 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 45 Resp: 400135
Ion Ratio Lower Upper
45 100
43 123.7 81.0 121.4#
87 25.1 20.3 30.5
59 11.4 9.0 13.4









#23
 Vinyl Acetate
 Concen: 274.498 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 1611770
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.8 11.8



#24
 1,1-Dichloroethane
 Concen: 54.560 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03

Tgt Ion: 63 Resp: 212751
 Ion Ratio Lower Upper
 63 100
 98 7.3 3.8 11.2
 100 4.2 2.3 6.9









#25
2-Butanone
Concen: 247.197 ug/l
RT: 4.538 min Scan# 566
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 329510
Ion Ratio Lower Upper
43 100
72 24.5 20.1 30.1



#26
2,2-Dichloropropane
Concen: 44.928 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 77 Resp: 145629
Ion Ratio Lower Upper
77 100
97 22.4 10.9 32.7





#27

cis-1,2-Dichloroethene

Concen: 53.677 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min


Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 129834

Ion Ratio Lower Upper

96 100

61 149.6 0.0 303.2

98 66.1 0.0 127.8



#28

Bromochloromethane

Concen: 53.668 ug/l

RT: 4.885 min Scan# 623

Delta R.T. -0.006 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 49 Resp: 91361


Ion Ratio Lower Upper

49 100

129 1.1 0.0 3.2

130 71.8 57.8 86.8





#29

Tetrahydrofuran

Concen: 260.441 ug/l

RT: 4.989 min Scan# 640

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 208846

Ion Ratio Lower Upper

42 100

72 42.8 34.2 51.2

71 39.4 31.9 47.9



#30

Chloroform

Concen: 54.443 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion: 83 Resp: 214531

Ion Ratio Lower Upper

83 100


85 61.7 51.8 77.6








#31
Cyclohexane
Concen: 53.387 ug/l
RT: 5.464 min Scan# 717
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 56 Resp: 173386




Ion	Ratio	Lower	Upper
56	100		
69	28.3	23.8	35.8
84	82.0	64.6	97.0



#32
1,1,1-Trichloroethane
Concen: 53.474 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 97 Resp: 178728

Ion	Ratio	Lower	Upper
97	100		
99	63.5	50.6	76.0
61	45.0	35.8	53.8





#33

1,2-Dichloroethane-d4

Concen: 53.282 ug/l

RT: 5.946 min Scan# 796

Delta R.T. 0.006 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 130953

Ion Ratio Lower Upper

65 100

67 52.1 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. 0.006 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 114 Resp: 278359

Ion Ratio Lower Upper

114 100

63 21.8 0.0 44.0

88 16.8 0.0 34.2





#35

Dibromofluoromethane

Concen: 53.721 ug/l

RT: 5.373 min Scan# 703

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:113 Resp: 100288

Ion Ratio Lower Upper

113 100

111 102.7 80.9 121.3

192 17.9 14.2 21.4



#36

1,1-Dichloropropene

Concen: 51.362 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 75 Resp: 140108


Ion Ratio Lower Upper

75 100

110 34.5 16.7 50.1

77 30.8 24.3 36.5









#37
Ethyl Acetate
Concen: 50.499 ug/l
RT: 4.696 min Scan# 591
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 147497
Ion Ratio Lower Upper
43 100
61 12.0 9.8 14.8
70 9.9 7.9 11.9



#38
Carbon Tetrachloride
Concen: 50.831 ug/l
RT: 5.666 min Scan# 751
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 117 Resp: 150656
Ion Ratio Lower Upper
117 100
119 97.5 73.4 110.2
121 31.1 23.8 35.6





#39

Methylcyclohexane

Concen: 52.363 ug/l

RT: 7.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 162143

Ion Ratio Lower Upper

83 100

55 82.9 66.4 99.6

98 49.7 38.1 57.1



#40

Benzene

Concen: 53.802 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03

Tgt Ion: 78 Resp: 445760

Ion Ratio Lower Upper

78 100


77 23.8 18.8 28.2







#41
Methacrylonitrile
Concen: 53.000 ug/l
RT: 4.903 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	41	Resp:	79917
Ion	Ratio	Lower	Upper
41	100		
39	55.8	43.3	64.9
67	73.8	58.6	87.8
52	32.5	26.0	39.0



#42
1,2-Dichloroethane
Concen: 51.951 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:	62	Resp:	163773
Ion	Ratio	Lower	Upper
62	100		
98	9.0	0.0	17.0








#43
Isopropyl Acetate
Concen: 51.410 ug/l
RT: 6.324 min Scan# 859
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 246265
Ion Ratio Lower Upper
43 100
61 19.2 15.2 22.8
87 12.3 9.6 14.4



#44
Trichloroethene
Concen: 56.204 ug/l
RT: 7.117 min Scan# 989
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:130 Resp: 112580
Ion Ratio Lower Upper
130 100
95 98.1 0.0 210.8





#45

1,2-Dichloropropane

Concen: 53.755 ug/l

RT: 7.415 min Scan# 1038

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 114044

Ion Ratio Lower Upper

63 100

65 31.6 24.7 37.1



#46

Dibromomethane

Concen: 53.014 ug/l

RT: 7.568 min Scan# 1063

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 93 Resp: 81845


Ion Ratio Lower Upper

93 100

95 83.7 65.0 97.6

174 91.9 71.6 107.4









#47
Bromodichloromethane
Concen: 53.260 ug/l
RT: 7.805 min Scan# 1102
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 169676
Ion Ratio Lower Upper
83 100
85 63.3 52.5 78.7
127 8.4 6.9 10.3



#48
Methyl methacrylate
Concen: 52.004 ug/l
RT: 7.677 min Scan# 1081
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 41 Resp: 124085
Ion Ratio Lower Upper
41 100
69 78.2 61.4 92.2
39 55.3 45.8 68.6









#49
1,4-Dioxane
Concen: 1102.850 ug/l
RT: 7.647 min Scan# 1076
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 88 Resp: 26594
Ion Ratio Lower Upper
88 100
43 42.3 30.2 45.2
58 74.2 59.5 89.3



#50
Toluene-d8
Concen: 53.308 ug/l
RT: 8.635 min Scan# 1238
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 98 Resp: 344346
Ion Ratio Lower Upper
98 100
100 67.4 53.0 79.4









#51
4-Methyl-2-Pentanone
Concen: 264.517 ug/l
RT: 8.555 min Scan# 1225
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 43 Resp: 702538
Ion Ratio Lower Upper
43 100
58 37.6 30.3 45.5



#52
Toluene
Concen: 53.368 ug/l
RT: 8.702 min Scan# 1249
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 92 Resp: 272421
Ion Ratio Lower Upper
92 100
91 169.4 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 51.527 ug/l

RT: 8.964 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 167442

Ion Ratio Lower Upper

75 100

77 32.1 25.4 38.2



#54

cis-1,3-Dichloropropene

Concen: 51.815 ug/l

RT: 8.354 min Scan# 1192

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 75 Resp: 179970


Ion Ratio Lower Upper

75 100

77 32.8 25.8 38.6

39 50.9 41.4 62.2





#55

1,1,2-Trichloroethane

Concen: 53.322 ug/l

RT: 9.134 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 104952





Ion Ratio Lower Upper

97 100

83 85.2 69.6 104.4

85 55.9 45.0 67.6

99 62.0 49.6 74.4



#56

Ethyl methacrylate

Concen: 53.586 ug/l

RT: 9.104 min Scan# 1315

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion: 69 Resp: 165642


Ion Ratio Lower Upper

69 100

41 71.6 58.6 87.8

39 38.5 31.5 47.3





#57

1,3-Dichloropropane

Concen: 54.129 ug/l

RT: 9.293 min Scan# 11346

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 183018

Ion Ratio Lower Upper

76 100

78 32.4 26.2 39.4



#58

2-Chloroethyl Vinyl ether

Concen: 259.650 ug/l

RT: 8.226 min Scan# 1171

Delta R.T. 0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion: 63 Resp: 384468

Ion Ratio Lower Upper

63 100

106 27.6 21.9 32.9









#59
2-Hexanone
Concen: 252.510 ug/l
RT: 9.415 min Scan# 1366
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 481662
Ion Ratio Lower Upper
43 100
58 52.8 26.1 78.2



#60
Dibromochloromethane
Concen: 53.465 ug/l
RT: 9.506 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:129 Resp: 121253
Ion Ratio Lower Upper
129 100
127 77.6 38.4 115.2








#61
1,2-Dibromoethane
Concen: 54.996 ug/l
RT: 9.592 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:107 Resp: 110255
Ion Ratio Lower Upper
107 100
109 95.0 75.3 112.9



#62
4-Bromofluorobenzene
Concen: 52.822 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03


Tgt Ion: 95 Resp: 129494
Ion Ratio Lower Upper
95 100
174 71.9 0.0 141.6
176 69.5 0.0 138.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 248998
Ion Ratio Lower Upper
117 100
82 58.7 47.8 71.6
119 32.7 25.2 37.8




#64
Tetrachloroethene
Concen: 52.205 ug/l
RT: 9.262 min Scan# 1341
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03



Tgt Ion:164 Resp: 84740
Ion Ratio Lower Upper
164 100
166 123.9 100.2 150.2
129 94.9 81.8 122.6
131 90.3 78.6 118.0





#65

Chlorobenzene

Concen: 52.793 ug/l

RT: 10.061 min Scan# 1472

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 298639

Ion Ratio Lower Upper

112 100

114 32.2 25.1 37.7



#66

1,1,1,2-Tetrachloroethane

Concen: 52.201 ug/l

RT: 10.146 min Scan# 1486

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion:131 Resp: 101896


Ion Ratio Lower Upper

131 100

133 98.0 47.3 142.0

119 65.0 32.0 96.0









#67
Ethyl Benzene
Concen: 52.982 ug/l
RT: 10.177 min Scan# 1491
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 517118
Ion Ratio Lower Upper
91 100
106 30.1 24.3 36.5



#68
m/p-Xylenes
Concen: 104.973 ug/l
RT: 10.287 min Scan# 1509
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:106 Resp: 381354
Ion Ratio Lower Upper
106 100
91 211.0 167.8 251.8









#69
o-Xylene
Concen: 53.332 ug/l
RT: 10.628 min Scan# 1564
Delta R.T. 0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:106 Resp: 187399
Ion Ratio Lower Upper
106 100
91 220.5 111.1 333.4



#70
Styrene
Concen: 53.259 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:104 Resp: 327913
Ion Ratio Lower Upper
104 100
78 52.5 43.1 64.7
103 53.9 43.1 64.7





#71

Bromoform

Concen: 51.882 ug/l

RT: 10.780 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX047687.D





Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173	Resp:	75620
Ion Ratio	Lower	Upper
173	100	
175	48.3	24.9 74.6
254	0.0	0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.006 min Scan# 1791


Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Tgt Ion:152	Resp:	120684
Ion Ratio	Lower	Upper
152	100	
115	86.4	44.9 134.5
150	175.9	0.0 352.0





#73

Isopropylbenzene

Concen: 51.835 ug/l

RT: 10.945 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 475596

Ion Ratio Lower Upper

105 100

120 25.5 12.8 38.4



#74

N-ethyl acetate

Concen: 51.785 ug/l

RT: 10.829 min Scan# 1598

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Tgt Ion: 43 Resp: 216669




Ion Ratio Lower Upper


43 100

70 40.2 32.0 48.0

55 23.3 18.9 28.3

61 23.4 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 52.421 ug/l

RT: 11.195 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 145518

Ion Ratio Lower Upper

83 100

131 9.8 5.0 14.9

85 64.8 31.6 95.0



#76

1,2,3-Trichloropropane

Concen: 64.800 ug/l

RT: 11.219 min Scan# 1662

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion: 75 Resp: 158337

Ion Ratio Lower Upper

75 100

77 30.9 18.8 56.4









#77
Bromobenzene
Concen: 51.947 ug/l
RT: 11.183 min Scan# 1656
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 117390
Ion Ratio Lower Upper
156 100
77 178.3 89.7 269.1
158 97.4 47.9 143.6



#78
n-propylbenzene
Concen: 52.147 ug/l
RT: 11.286 min Scan# 1673
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 91 Resp: 565595
Ion Ratio Lower Upper
91 100
120 21.9 11.1 33.1





#79

2-Chlorotoluene

Concen: 51.361 ug/l

RT: 11.347 min Scan# 1683

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 341195

Ion Ratio Lower Upper

91 100

126 32.4 16.1 48.3



#80

1,3,5-Trimethylbenzene

Concen: 51.564 ug/l

RT: 11.433 min Scan# 1697

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion: 105 Resp: 388702

Ion Ratio Lower Upper

105 100

120 48.1 23.6 70.8









#81
 trans-1,4-Dichloro-2-butene
 Concen: 45.792 ug/l
 RT: 11.000 min Scan# 1626
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 43877
 Ion Ratio Lower Upper
 75 100
 53 108.1 83.4 125.2
 89 47.0 36.3 54.5



#82
 4-Chlorotoluene
 Concen: 51.132 ug/l
 RT: 11.439 min Scan# 1698
 Delta R.T. -0.000 min
 Lab File: VX047687.D
 Acq: 19 Sep 2025 19:03

Tgt Ion: 91 Resp: 401172
 Ion Ratio Lower Upper
 91 100
 126 29.5 14.2 42.6





#83

tert-Butylbenzene

Concen: 50.797 ug/l

RT: 11.695 min Scan# 1740

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 391982

Ion Ratio Lower Upper

119 100

91 58.7 29.6 88.8

134 22.6 11.0 33.0



#84

1,2,4-Trimethylbenzene

Concen: 52.011 ug/l

RT: 11.732 min Scan# 1746

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion:105 Resp: 397498

Ion Ratio Lower Upper

105 100

120 43.8 22.1 66.1









#85
sec-Butylbenzene
Concen: 51.055 ug/l
RT: 11.872 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 467967
Ion Ratio Lower Upper
105 100
134 19.6 9.8 29.5



#86
p-Isopropyltoluene
Concen: 50.599 ug/l
RT: 11.988 min Scan# 1788
Delta R.T. -0.006 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:119 Resp: 392640
Ion Ratio Lower Upper
119 100
134 25.8 12.6 37.8
91 24.8 12.6 37.6





#87

1,3-Dichlorobenzene

Concen: 51.456 ug/l

RT: 11.951 min Scan# 1782

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 215979

Ion Ratio Lower Upper

146 100

111 42.5 21.5 64.5

148 62.8 32.0 96.0



#88

1,4-Dichlorobenzene

Concen: 50.659 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion:146 Resp: 218234


Ion Ratio Lower Upper

146 100

111 41.3 20.8 62.4

148 63.3 31.7 95.1





#89

n-Butylbenzene

Concen: 50.892 ug/l

RT: 12.311 min Scan# 1841

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 365418

Ion Ratio Lower Upper

91 100

92 53.6 27.6 82.7

134 25.0 12.6 37.8



#90

Hexachloroethane

Concen: 50.250 ug/l

RT: 12.518 min Scan# 1875

Delta R.T. -0.000 min

Lab File: VX047687.D




Acq: 19 Sep 2025 19:03


Tgt Ion: 117 Resp: 68460

Ion Ratio Lower Upper

117 100

201 66.7 34.2 102.5









#91
1,2-Dichlorobenzene
Concen: 51.610 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03




Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:146 Resp: 206378
Ion Ratio Lower Upper
146 100
111 43.8 22.2 66.6
148 64.4 31.8 95.4



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.873 ug/l
RT: 12.920 min Scan# 1941
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion: 75 Resp: 26906
Ion Ratio Lower Upper
75 100
155 80.6 37.7 113.1
157 104.0 49.8 149.4





#93

1,2,4-Trichlorobenzene

Concen: 50.815 ug/l

RT: 13.567 min Scan# 2047

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 132061

Ion Ratio Lower Upper

180 100

182 93.5 46.6 139.8

145 32.1 16.4 49.1



#94

Hexachlorobutadiene

Concen: 46.163 ug/l

RT: 13.707 min Scan# 2070

Delta R.T. -0.000 min

Lab File: VX047687.D

Acq: 19 Sep 2025 19:03




Tgt Ion:225 Resp: 40751


Ion Ratio Lower Upper

225 100

223 65.3 31.7 95.1


227 64.6 32.0 96.0








#95
Naphthalene
Concen: 49.995 ug/l
RT: 13.756 min Scan# 2078
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 395649
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.2
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 49.173 ug/l
RT: 13.938 min Scan# 2108
Delta R.T. -0.000 min
Lab File: VX047687.D
Acq: 19 Sep 2025 19:03

Tgt Ion:180 Resp: 118876
Ion Ratio Lower Upper
180 100
182 98.1 47.6 142.8
145 34.9 17.9 53.7

