

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047662.D
 Acq On : 19 Sep 2025 10:11
 Operator : JC/MD
 Sample : VX0919WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0919WBL01

Quant Time: Sep 20 00:26:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	194256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	363806	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	339161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	175228	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	178788	57.822	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.640%
35) Dibromofluoromethane	5.373	113	128166	52.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.060%
50) Toluene-d8	8.634	98	441682	52.317	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.640%
62) 4-Bromofluorobenzene	11.061	95	180450	56.319	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047662.D
Acq On : 19 Sep 2025 10:11
Operator : JC/MD
Sample : VX0919WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0919WBL01

Quant Time: Sep 20 00:26:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

