

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047666.D
 Acq On : 19 Sep 2025 11:47
 Operator : JC/MD
 Sample : Q3095-25DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20250910DL

Quant Time: Sep 20 00:29:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	189981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	344496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	309572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	151017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	164845	54.512	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.020%	
35) Dibromofluoromethane	5.379	113	124924	54.071	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.140%	
50) Toluene-d8	8.634	98	434109	54.302	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.600%	
62) 4-Bromofluorobenzene	11.067	95	167474	55.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.400%	
Target Compounds						
44) Trichloroethene	7.110	130	144090	58.124	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047666.D
Acq On : 19 Sep 2025 11:47
Operator : JC/MD
Sample : Q3095-25DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20250910DL

Quant Time: Sep 20 00:29:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

