

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047669.D
 Acq On : 19 Sep 2025 12:50
 Operator : JC/MD
 Sample : Q3135-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121

Quant Time: Sep 20 00:30:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	179125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	335824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	306398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	149153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	159501	55.941	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.880%
35) Dibromofluoromethane	5.379	113	117460	52.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.300%
50) Toluene-d8	8.635	98	405117	51.984	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.960%
62) 4-Bromofluorobenzene	11.061	95	155822	52.685	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.360%
Target Compounds						
16) Acetone	2.380	43	2635	2.125	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047669.D
Acq On : 19 Sep 2025 12:50
Operator : JC/MD
Sample : Q3135-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121

Quant Time: Sep 20 00:30:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

