

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047674.D
 Acq On : 19 Sep 2025 14:34
 Operator : JC/MD
 Sample : Q3095-32
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20250911

Quant Time: Sep 20 00:32:29 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	188618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	342292	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	316267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	150121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	153541	51.141	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.280%	
35) Dibromofluoromethane	5.373	113	115022	50.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.220%	
50) Toluene-d8	8.634	98	395238	49.758	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.520%	
62) 4-Bromofluorobenzene	11.061	95	151815	50.360	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.720%	
Target Compounds						
					Qvalue	
44) Trichloroethene	7.116	130	4544	1.845	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047674.D
Acq On : 19 Sep 2025 14:34
Operator : JC/MD
Sample : Q3095-32
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20250911

Quant Time: Sep 20 00:32:29 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

