

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047678.D
 Acq On : 19 Sep 2025 15:57
 Operator : JC/MD
 Sample : Q3097-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20250911

Quant Time: Sep 20 00:34:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	166867	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	308997	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	279494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	147555	55.553	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	111.100%	
35) Dibromofluoromethane	5.379	113	109111	52.652	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	105.300%	
50) Toluene-d8	8.635	98	379125	52.872	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	105.740%	
62) 4-Bromofluorobenzene	11.061	95	145340	53.407	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	106.820%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	2520	1.306	ug/l	97
16) Acetone	2.373	43	2476	2.144	ug/l	90
44) Trichloroethene	7.110	130	187117	84.153	ug/l	99
64) Tetrachloroethene	9.262	164	779	0.428	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047678.D
Acq On : 19 Sep 2025 15:57
Operator : JC/MD
Sample : Q3097-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20250911

Quant Time: Sep 20 00:34:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

