

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047679.D
 Acq On : 19 Sep 2025 16:17
 Operator : JC/MD
 Sample : Q3097-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-20250911

Quant Time: Sep 20 00:34:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	180253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	329994	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	279542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	137704	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	157402	54.860	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 109.720%		
35) Dibromofluoromethane	5.379	113	118552	53.568	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 107.140%		
50) Toluene-d8	8.635	98	382375	49.933	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 99.860%		
62) 4-Bromofluorobenzene	11.061	95	149571	51.465	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 102.920%		
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.343	101	4559	2.187	ug/l	98
12) 1,1-Dichloroethene	2.331	96	5262	2.457	ug/l	96
24) 1,1-Dichloroethane	3.611	63	7030	1.544	ug/l	94
27) cis-1,2-Dichloroethene	4.489	96	7496	2.655	ug/l	89
30) Chloroform	5.080	83	4859	1.056	ug/l #	77
38) Carbon Tetrachloride	5.666	117	2098	0.597	ug/l #	84
44) Trichloroethene	7.111	130	3014300	1269.373	ug/l	96
64) Tetrachloroethene	9.263	164	4744	2.603	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047679.D
Acq On : 19 Sep 2025 16:17
Operator : JC/MD
Sample : Q3097-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-20250911

Quant Time: Sep 20 00:34:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

