

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047681.D
 Acq On : 19 Sep 2025 16:59
 Operator : JC/MD
 Sample : Q3097-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20250911

Quant Time: Sep 20 00:35:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	172130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	316059	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	289018	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	139300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	154151	56.262	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.520%
35) Dibromofluoromethane	5.373	113	115726	54.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.200%
50) Toluene-d8	8.634	98	402399	54.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.720%
62) 4-Bromofluorobenzene	11.067	95	153502	55.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.300%
Target Compounds						
						Qvalue
44) Trichloroethene	7.116	130	62687	27.562	ug/l	92
64) Tetrachloroethene	9.262	164	3487	1.851	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047681.D
Acq On : 19 Sep 2025 16:59
Operator : JC/MD
Sample : Q3097-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20250911

Quant Time: Sep 20 00:35:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

