

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047683.D
 Acq On : 19 Sep 2025 17:40
 Operator : JC/MD
 Sample : Q3097-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20250911

Quant Time: Sep 20 00:36:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	171618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	317965	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	289328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	137612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	154083	56.405	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.800%
35) Dibromofluoromethane	5.373	113	116291	54.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.060%
50) Toluene-d8	8.635	98	401515	54.415	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.840%
62) 4-Bromofluorobenzene	11.061	95	154347	55.117	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.240%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	10394	5.236	ug/l	100
12) 1,1-Dichloroethene	2.325	96	1056	0.518	ug/l #	87
16) Acetone	2.380	43	1714	1.443	ug/l #	88
44) Trichloroethene	7.117	130	93951	41.061	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047683.D
Acq On : 19 Sep 2025 17:40
Operator : JC/MD
Sample : Q3097-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20250911

Quant Time: Sep 20 00:36:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

