

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047685.D
 Acq On : 19 Sep 2025 18:22
 Operator : JC/MD
 Sample : Q3097-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20250911

Quant Time: Sep 20 00:37:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	181407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	335366	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	298999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	146680	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	144009	49.873	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.740%		
35) Dibromofluoromethane	5.373	113	110247	49.017	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.040%		
50) Toluene-d8	8.634	98	372718	47.892	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.780%		
62) 4-Bromofluorobenzene	11.067	95	144318	48.862	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.720%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	3451	1.645	ug/l	92
12) 1,1-Dichloroethene	2.331	96	1183	0.549	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	4591	0.584	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	4109	1.446	ug/l	97
44) Trichloroethene	7.110	130	337410	139.813	ug/l	99
64) Tetrachloroethene	9.256	164	17507	8.982	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047685.D
Acq On : 19 Sep 2025 18:22
Operator : JC/MD
Sample : Q3097-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20250911

Quant Time: Sep 20 00:37:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

