

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047660.D
 Acq On : 19 Sep 2025 09:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 20 00:25:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	-0.01
2 T	Dichlorodifluoromethane	0.518	0.650	-25.5#	126	0.00
3 P	Chloromethane	0.680	0.796	-17.1	117	0.00
4 C	Vinyl Chloride	0.666	0.765	-14.9#	112	0.00
5 T	Bromomethane	0.422	0.509	-20.6	116	0.00
6 T	Chloroethane	0.445	0.487	-9.4	107	0.00
7 T	Trichlorofluoromethane	0.989	1.055	-6.7	103	0.00
8 T	Diethyl Ether	0.405	0.431	-6.4	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.606	-4.8	101	0.00
10 T	Methyl Iodide	0.902	0.957	-6.1	102	0.00
11 T	Tert butyl alcohol	0.089	0.091	-2.2	99	0.00
12 CM	1,1-Dichloroethene	0.594	0.636	-7.1#	102	0.00
13 T	Acrolein	0.082	0.091	-11.0	108	0.00
14 T	Allyl chloride	1.226	1.277	-4.2	102	0.00
15 T	Acrylonitrile	0.328	0.368	-12.2	103	0.00
16 T	Acetone	0.346	0.370	-6.9	113	0.00
17 T	Carbon Disulfide	1.626	1.859	-14.3	115	0.00
18 T	Methyl Acetate	0.770	0.823	-6.9	91	0.00
19 T	Methyl tert-butyl Ether	2.169	2.275	-4.9	98	0.00
20 T	Methylene Chloride	0.732	0.763	-4.2	102	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.690	-7.8	103	0.00
22 T	Diisopropyl ether	2.376	2.524	-6.2	99	-0.01
23 T	Vinyl Acetate	1.901	2.110	-11.0	103	0.00
24 P	1,1-Dichloroethane	1.263	1.331	-5.4	100	0.00
25 T	2-Butanone	0.432	0.467	-8.1	103	0.00
26 T	2,2-Dichloropropane	1.050	1.086	-3.4	99	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.828	-5.7	99	0.00
28 T	Bromochloromethane	0.551	0.612	-11.1	103	-0.01
29 T	Tetrahydrofuran	0.260	0.286	-10.0	101	0.00
30 C	Chloroform	1.276	1.333	-4.5#	98	0.00
31 T	Cyclohexane	1.052	1.072	-1.9	101	0.00
32 T	1,1,1-Trichloroethane	1.082	1.113	-2.9	96	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.779	2.1	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.335	0.329	1.8	99	0.00
36 T	1,1-Dichloropropene	0.490	0.485	1.0	99	-0.01
37 T	Ethyl Acetate	0.525	0.547	-4.2	103	-0.01
38 T	Carbon Tetrachloride	0.532	0.515	3.2	95	-0.01
39 T	Methylcyclohexane	0.556	0.565	-1.6	100	0.00
40 TM	Benzene	1.488	1.532	-3.0	100	0.00
41 T	Methacrylonitrile	0.271	0.293	-8.1	102	0.00
42 TM	1,2-Dichloroethane	0.566	0.567	-0.2	97	0.00
43 T	Isopropyl Acetate	0.860	0.880	-2.3	98	0.00
44 TM	Trichloroethene	0.360	0.371	-3.1	100	0.00
45 C	1,2-Dichloropropane	0.381	0.386	-1.3#	97	0.00
46 T	Dibromomethane	0.277	0.285	-2.9	99	0.00
47 T	Bromodichloromethane	0.572	0.580	-1.4	95	0.00
48 T	Methyl methacrylate	0.429	0.442	-3.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
 Data File : VX047660.D
 Acq On : 19 Sep 2025 09:22
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 20 00:25:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.005	-25.0	114	0.00
50 S	Toluene-d8	1.160	1.122	3.3	98	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.504	-5.7	98	0.00
52 CM	Toluene	0.917	0.930	-1.4#	99	0.00
53 T	t-1,3-Dichloropropene	0.584	0.600	-2.7	97	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.646	-3.5	99	0.00
55 T	1,1,2-Trichloroethane	0.354	0.359	-1.4	98	0.00
56 T	Ethyl methacrylate	0.555	0.590	-6.3	98	0.00
57 T	1,3-Dichloropropane	0.607	0.632	-4.1	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.278	-22.5	102	0.00
59 T	2-Hexanone	0.343	0.360	-5.0	100	0.00
60 T	Dibromochloromethane	0.407	0.423	-3.9	97	0.00
61 T	1,2-Dibromoethane	0.360	0.383	-6.4	101	0.00
62 S	4-Bromofluorobenzene	0.440	0.425	3.4	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.326	0.319	2.1	99	0.00
65 PM	Chlorobenzene	1.136	1.134	0.2	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.393	-0.3	96	0.00
67 C	Ethyl Benzene	1.960	1.969	-0.5#	97	0.00
68 T	m/p-Xylenes	0.729	0.735	-0.8	97	0.00
69 T	o-Xylene	0.706	0.714	-1.1	96	0.00
70 T	Styrene	1.236	1.272	-2.9	99	0.00
71 P	Bromoform	0.293	0.298	-1.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.801	3.802	-0.0	95	0.00
74 T	N-amyl acetate	1.733	1.810	-4.4	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.194	-3.8	99	0.00
76 T	1,2,3-Trichloropropane	1.012	0.965	4.6	87	0.00
77 T	Bromobenzene	0.936	0.939	-0.3	97	0.00
78 T	n-propylbenzene	4.494	4.495	-0.0	95	0.00
79 T	2-Chlorotoluene	2.752	2.745	0.3	96	0.00
80 T	1,3,5-Trimethylbenzene	3.123	3.156	-1.1	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.400	-0.8	98	0.00
82 T	4-Chlorotoluene	3.251	3.246	0.2	96	0.00
83 T	tert-Butylbenzene	3.197	3.082	3.6	93	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.199	-1.0	97	0.00
85 T	sec-Butylbenzene	3.797	3.769	0.7	95	0.00
86 T	p-Isopropyltoluene	3.215	3.199	0.5	95	0.00
87 T	1,3-Dichlorobenzene	1.739	1.744	-0.3	98	0.00
88 T	1,4-Dichlorobenzene	1.785	1.754	1.7	97	0.00
89 T	n-Butylbenzene	2.975	2.976	-0.0	96	0.00
90 T	Hexachloroethane	0.564	0.563	0.2	96	0.00
91 T	1,2-Dichlorobenzene	1.657	1.660	-0.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.242	-3.9	97	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.091	-1.3	96	0.00
94 T	Hexachlorobutadiene	0.366	0.365	0.3	97	0.00
95 T	Naphthalene	3.279	3.303	-0.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX091925\
Data File : VX047660.D
Acq On : 19 Sep 2025 09:22
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 20 00:25:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.028	-2.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6