

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012551.D
 Acq On : 20 Sep 2019 12:52
 Operator : JC/SP
 Sample : K4858-22DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20190910DL

Quant Time: Sep 21 01:49:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	185734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	293131	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244601	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98320	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	123172	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
35) Dibromofluoromethane	5.49	113	88337	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
50) Toluene-d8	8.71	98	340231	51.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.62%	
62) 4-Bromofluorobenzene	11.13	95	119696	43.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.80%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	41060	28.050	ug/l	97
44) Trichloroethene	7.21	130	7218	4.844	ug/l	96
64) Tetrachloroethene	9.33	164	2652	2.366	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012551.D
Acq On : 20 Sep 2019 12:52
Operator : JC/SP
Sample : K4858-22DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20190910DL

Quant Time: Sep 21 01:49:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

