

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012556.D
 Acq On : 20 Sep 2019 14:48
 Operator : JC/SP
 Sample : K4883-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW2-R2-1

Quant Time: Sep 21 02:10:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	181864	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	292493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93914	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122034	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
35) Dibromofluoromethane	5.49	113	86609	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	8.71	98	331351	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
62) 4-Bromofluorobenzene	11.14	95	116207	42.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.42%	
Target Compounds						
16) Acetone	2.43	43	12574	2.683	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012556.D
Acq On : 20 Sep 2019 14:48
Operator : JC/SP
Sample : K4883-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW2-R2-1

Quant Time: Sep 21 02:10:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

