

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012576.D
 Acq On : 20 Sep 2019 23:22
 Operator : JC/SP
 Sample : K4930-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190916

Quant Time: Sep 21 03:51:44 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	278461	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	90729	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	115470	45.85	ug/l	0.00
Spiked Amount			Recovery	=	91.70%	
35) Dibromofluoromethane	5.48	113	83734	49.68	ug/l	0.00
Spiked Amount			Recovery	=	99.36%	
50) Toluene-d8	8.71	98	321665	51.56	ug/l	0.00
Spiked Amount			Recovery	=	103.12%	
62) 4-Bromofluorobenzene	11.14	95	112157	43.30	ug/l	0.00
Spiked Amount			Recovery	=	86.60%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012576.D
Acq On : 20 Sep 2019 23:22
Operator : JC/SP
Sample : K4930-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190916

Quant Time: Sep 21 03:51:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

