

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012580.D
 Acq On : 21 Sep 2019 00:56
 Operator : JC/SP
 Sample : K4956-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0547

Quant Time: Sep 21 04:00:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178126	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291009	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	245377	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	100620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119294	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
35) Dibromofluoromethane	5.48	113	86870	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
50) Toluene-d8	8.71	98	330930	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.14	95	119225	44.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.08%	
Target Compounds						
11) Tert butyl alcohol	3.03	59	5606	2.854	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	1078	0.621	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012580.D
Acq On : 21 Sep 2019 00:56
Operator : JC/SP
Sample : K4956-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0547

Quant Time: Sep 21 04:00:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

