

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012584.D
 Acq On : 21 Sep 2019 02:29
 Operator : JC/SP
 Sample : K4930-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS037-190916

Quant Time: Sep 21 04:31:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	282757	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	95401	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	118633	46.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.54%	
35) Dibromofluoromethane	5.49	113	85302	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
50) Toluene-d8	8.71	98	323781	51.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.14	95	114066	43.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.74%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	10919	6.258	ug/l	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012584.D
Acq On : 21 Sep 2019 02:29
Operator : JC/SP
Sample : K4930-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS037-190916

Quant Time: Sep 21 04:31:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

