

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012586.D
 Acq On : 21 Sep 2019 03:16
 Operator : JC/SP
 Sample : K4930-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS036-190916

Quant Time: Sep 21 04:34:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	176375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	279375	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231968	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	89324	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	117395	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
35) Dibromofluoromethane	5.49	113	86099	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
50) Toluene-d8	8.71	98	320052	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
62) 4-Bromofluorobenzene	11.14	95	107790	41.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.96%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.59	96	48208	28.025	ug/l	85
44) Trichloroethene	7.21	130	1816	1.279	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012586.D
Acq On : 21 Sep 2019 03:16
Operator : JC/SP
Sample : K4930-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS036-190916

Quant Time: Sep 21 04:34:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

