

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX092019\
 Data File : VX012590.D
 Acq On : 21 Sep 2019 04:49
 Operator : JC/SP
 Sample : K4978-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-0164

Quant Time: Sep 21 06:46:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	181687	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	288694	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	241852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101813	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117465	45.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.22%	
35) Dibromofluoromethane	5.49	113	84237	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
50) Toluene-d8	8.71	98	328617	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.14	95	118499	44.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.26%	
Target Compounds						
					Qvalue	
40) Benzene	6.14	78	2257	0.405	ug/l	90
67) Ethyl Benzene	10.25	91	38659	5.637	ug/l	100
68) m/p-Xylenes	10.36	106	1883	0.755	ug/l	91
73) Isopropylbenzene	11.01	105	29066	4.550	ug/l	99
78) n-propylbenzene	11.35	91	30663	4.307	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	1940	0.347	ug/l	94
85) sec-Butylbenzene	11.94	105	5426	0.843	ug/l #	76
95) Naphthalene	13.83	128	4199	0.598	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012590.D
Acq On : 21 Sep 2019 04:49
Operator : JC/SP
Sample : K4978-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-0164

Quant Time: Sep 21 06:46:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

