

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX092019\
 Data File : VX012591.D
 Acq On : 21 Sep 2019 05:12
 Operator : JC/SP
 Sample : K4978-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-0165

Quant Time: Sep 21 06:55:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 17 15:27:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	182964	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	291478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	241818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	96532	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	121305	46.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.52%	
35) Dibromofluoromethane	5.48	113	87362	49.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.04%	
50) Toluene-d8	8.71	98	331384	50.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.50%	
62) 4-Bromofluorobenzene	11.14	95	115561	42.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.24%	
Target Compounds						
31) Cyclohexane	5.57	56	2771	1.530	ug/l #	74
40) Benzene	6.14	78	40106	7.120	ug/l	95
73) Isopropylbenzene	11.01	105	4368	0.721	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX092019\
Data File : VX012591.D
Acq On : 21 Sep 2019 05:12
Operator : JC/SP
Sample : K4978-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-0165

Quant Time: Sep 21 06:55:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X091719W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 17 15:27:10 2019
Response via : Initial Calibration

