

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
 Data File : VX024353.D
 Acq On : 20 Sep 2021 16:05
 Operator : JC/MD
 Sample : M3817-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-02-(0.3)

Quant Time: Sep 21 08:41:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148053	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259953	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	224289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	107799	51.061	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.120%
35) Dibromofluoromethane	5.397	113	86923	49.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.720%
50) Toluene-d8	8.652	98	269559	42.444	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	84.880%#
62) 4-Bromofluorobenzene	11.085	95	96111	38.886	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.780%#
Target Compounds						
						Qvalue
16) Acetone	2.391	43	3465	3.849	ug/l #	84
20) Methylene Chloride	2.794	84	1227	0.710	ug/l #	84
30) Chloroform	5.092	83	1065	0.350	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
Data File : VX024353.D
Acq On : 20 Sep 2021 16:05
Operator : JC/MD
Sample : M3817-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-CB-02-(0.3)

Quant Time: Sep 21 08:41:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

