

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
 Data File : VX024360.D
 Acq On : 20 Sep 2021 18:50
 Operator : JC/MD
 Sample : M3818-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-860-862

Quant Time: Sep 21 08:43:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	151581	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264351	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	215554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92651	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111579	51.621	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.240%	
35) Dibromofluoromethane	5.397	113	88820	50.100	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.200%	
50) Toluene-d8	8.653	98	275485	42.655	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 85.320%#	
62) 4-Bromofluorobenzene	11.085	95	99165	39.455	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 78.900%#	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7631	8.280	ug/l	99
20) Methylene Chloride	2.800	84	953	0.538	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
Data File : VX024360.D
Acq On : 20 Sep 2021 18:50
Operator : JC/MD
Sample : M3818-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-860-862

Quant Time: Sep 21 08:43:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

