

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
 Data File : VX024361.D
 Acq On : 20 Sep 2021 19:14
 Operator : JC/MD
 Sample : M3818-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-885-887

Quant Time: Sep 21 08:43:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	226920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103255	52.740	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.480%	
35) Dibromofluoromethane	5.397	113	81100	51.509	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.020%	
50) Toluene-d8	8.653	98	296185	51.638	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.280%	
62) 4-Bromofluorobenzene	11.085	95	105143	47.103	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.200%	
Target Compounds						Qvalue
11) Tert butyl alcohol	2.947	59	13411	12.391	ug/l #	80
16) Acetone	2.386	43	25567	30.627	ug/l	95
20) Methylene Chloride	2.794	84	1714	1.069	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
Data File : VX024361.D
Acq On : 20 Sep 2021 19:14
Operator : JC/MD
Sample : M3818-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-885-887

Quant Time: Sep 21 08:43:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

