

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
 Data File : VX024368.D
 Acq On : 20 Sep 2021 23:32
 Operator : JC/MD
 Sample : VX0920WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0920WBL02

Quant Time: Sep 21 08:44:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	138166	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	219311	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93503	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	97185	49.327	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.660%
35) Dibromofluoromethane	5.397	113	82868	56.342	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.680%
50) Toluene-d8	8.653	98	272528	50.863	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.720%
62) 4-Bromofluorobenzene	11.085	95	102600	49.205	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
Data File : VX024368.D
Acq On : 20 Sep 2021 23:32
Operator : JC/MD
Sample : VX0920WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0920WBL02

Quant Time: Sep 21 08:44:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

