

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
 Data File : VX024372.D
 Acq On : 21 Sep 2021 01:06
 Operator : JC/MD
 Sample : M3782-04DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20210914DL

Quant Time: Sep 21 08:45:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 14 12:18:04 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	255369	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94716	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	106961	51.789	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.580%
35) Dibromofluoromethane	5.397	113	81805	47.766	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.540%
50) Toluene-d8	8.653	98	269902	43.261	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	86.520%#
62) 4-Bromofluorobenzene	11.085	95	112536	46.349	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.700%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	1566	1.098	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	969	0.574	ug/l	98
44) Trichloroethene	7.135	130	85492	45.185	ug/l	98
64) Tetrachloroethene	9.281	164	2085	1.086	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092021\
Data File : VX024372.D
Acq On : 21 Sep 2021 01:06
Operator : JC/MD
Sample : M3782-04DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20210914DL

Quant Time: Sep 21 08:45:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 14 12:18:04 2021
Response via : Initial Calibration

