

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037820.D
 Acq On : 20 Sep 2023 10:02
 Operator : JC/MD
 Sample : VX0920MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0920MBL01

Quant Time: Sep 21 05:51:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	247700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	410566	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	390174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	196365	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	149994	46.090	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.180%
35) Dibromofluoromethane	5.385	113	130726	47.587	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.646	98	505107	49.485	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.079	95	182522	49.209	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037820.D
Acq On : 20 Sep 2023 10:02
Operator : JC/MD
Sample : VX0920MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0920MBL01

Quant Time: Sep 21 05:51:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

