

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037821.D
 Acq On : 20 Sep 2023 10:25
 Operator : JC/MD
 Sample : VX0920WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0920WBL01

Quant Time: Sep 21 05:52:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	287740	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	481854	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	461021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	236259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174752	46.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.460%
35) Dibromofluoromethane	5.385	113	152429	47.278	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.560%
50) Toluene-d8	8.647	98	576052	48.086	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.079	95	218088	50.099	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037821.D
Acq On : 20 Sep 2023 10:25
Operator : JC/MD
Sample : VX0920WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0920WBL01

Quant Time: Sep 21 05:52:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

