

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037827.D
 Acq On : 20 Sep 2023 12:47
 Operator : JC/MD
 Sample : 04526-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-TB-20230914

Quant Time: Sep 21 05:55:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	280493	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	471971	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	448358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	232074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171741	46.602	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.200%
35) Dibromofluoromethane	5.391	113	149988	47.495	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.980%
50) Toluene-d8	8.653	98	563154	47.994	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.980%
62) 4-Bromofluorobenzene	11.079	95	213814	50.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037827.D
Acq On : 20 Sep 2023 12:47
Operator : JC/MD
Sample : 04526-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-TB-20230914

Quant Time: Sep 21 05:55:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

