

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037830.D
 Acq On : 20 Sep 2023 13:56
 Operator : JC/MD
 Sample : 04526-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-360-362

Quant Time: Sep 21 05:56:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	268468	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	451599	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	434104	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	224685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	165853	47.020	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.040%
35) Dibromofluoromethane	5.391	113	142215	47.065	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.120%
50) Toluene-d8	8.647	98	545776	48.611	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	205108	50.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037830.D
Acq On : 20 Sep 2023 13:56
Operator : JC/MD
Sample : 04526-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-360-362

Quant Time: Sep 21 05:56:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

