

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037831.D
 Acq On : 20 Sep 2023 14:20
 Operator : JC/MD
 Sample : 04526-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-380-382

Quant Time: Sep 21 05:56:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	263545	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	442199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420586	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	218232	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	164396	47.478	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.960%
35) Dibromofluoromethane	5.391	113	140726	47.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	8.653	98	533327	48.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.020%
62) 4-Bromofluorobenzene	11.079	95	197862	49.529	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037831.D
Acq On : 20 Sep 2023 14:20
Operator : JC/MD
Sample : 04526-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-380-382

Quant Time: Sep 21 05:56:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

