

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
 Data File : VX037833.D
 Acq On : 20 Sep 2023 15:06
 Operator : JC/MD
 Sample : 04526-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW10A-GW-420-422

Quant Time: Sep 21 05:57:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	258535	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	434984	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	417236	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	219074	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161788	47.630	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.260%
35) Dibromofluoromethane	5.385	113	138156	47.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.940%
50) Toluene-d8	8.653	98	528991	48.916	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	201012	51.152	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.300%
Target Compounds						
16) Acetone	2.380	43	4305	2.635	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092023\
Data File : VX037833.D
Acq On : 20 Sep 2023 15:06
Operator : JC/MD
Sample : 04526-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW10A-GW-420-422

Quant Time: Sep 21 05:57:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

