

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024381.D
 Acq On : 21 Sep 2021 11:33
 Operator : JC/MD
 Sample : VX0921WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0921WBL01

Quant Time: Sep 21 15:39:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	160140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	281413	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271507	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	118152	51.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.480%
35) Dibromofluoromethane	5.397	113	92867	49.207	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	8.653	98	324834	47.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.500%
62) 4-Bromofluorobenzene	11.085	95	130347	48.717	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024381.D
Acq On : 21 Sep 2021 11:33
Operator : JC/MD
Sample : VX0921WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0921WBL01

Quant Time: Sep 21 15:39:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

