

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024386.D
 Acq On : 21 Sep 2021 15:14
 Operator : JC/MD
 Sample : M3782-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20210914

Quant Time: Sep 21 15:45:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	151881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	267330	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	244527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108927	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	108029	49.880	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.760%		
35) Dibromofluoromethane	5.397	113	89772	50.073	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.140%		
50) Toluene-d8	8.653	98	318457	48.759	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.520%		
62) 4-Bromofluorobenzene	11.085	95	114211	44.935	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.860%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	2498	1.711	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.331	101	17674	11.820	ug/l	95
10) Methyl Iodide	2.471	142	1245	0.602	ug/l	99
12) 1,1-Dichloroethene	2.319	96	3551	2.536	ug/l	85
16) Acetone	2.386	43	3608	3.907	ug/l	89
27) cis-1,2-Dichloroethene	4.501	96	3947	2.229	ug/l	90
30) Chloroform	5.105	83	3921	1.257	ug/l	86
38) Carbon Tetrachloride	5.690	117	1508	0.545	ug/l	92
44) Trichloroethene	7.135	130	268074	135.346	ug/l	99
49) 1,4-Dioxane	7.671	88	1076	22.592	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024386.D
Acq On : 21 Sep 2021 15:14
Operator : JC/MD
Sample : M3782-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20210914

Quant Time: Sep 21 15:45:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

